

# MO PCR – quick set-up guide

Plug-in the 16 tube heating block (A,B) using power adapter (C) and 4 point connection. Power-up tablet / mobile phone, select MO App (E) & establish Bluetooth connection (F). Select last protocol (G) & what should be the only file on there (RFP2024), check conditions match protocol (H). Once samples are in the block you can hit RUN and you will hear the fan start to whir ... 71 minutes later it will be gel time

